AMENDMENT NO. 1 MARCH 2022

TO

IS 14709 : 1999 *N*-BUTYL ACRYLATE — SPECIFICATION

(*Page* 2, *Annex* A, *clause* **A-2.2**) — Substitute the following NOTE for the existing NOTE after clause **A-2.2**.

'NOTE — The above gas chromatographic conditions are suggestive. However, any GC having different columns (packed/capillary having different length/diameter/film thickness) and different carrier gas (He, H_2 or N_2), with different calibration technique (internal standard, external standard, area normalization) may be used provided standardization/ calibrations are done after setting up chromatographic conditions for required resolution.'

(PCD 09)

Publication Unit, BIS, New Delhi, India